

Certificate of Performance

Phos 1 Reagent



Product Code 984644
984646
Lot Number 8082

Date 02.05.2008


Expiry Date 30.04.2009

Reagent	Lot Number
Phos 1	8082
Phos 2	8083
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.003
0.5	P20 02 08	0.152
1	P20 02 08	0.286
1.5	P20 02 08	0.445
2	P20 02 08	0.577

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.004	-0.005 to 0.010	Pass
Sensitivity	Abs/ppm 0.288	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
*Stable on Board 5 days
Barcode ID = 763
* Based on results obtained in our Research Facility

Dr. T. Russell