

Certificate of Performance

Sulphate Reagent



Product Code 984648 20 ml vial
984649 60 ml vial
Lot Number 8143


Date 17.07.08 Expiry Date 15.07.09

Reagent	Lot Number
Sulphate	8143
Test Method	LTP01
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.002
250	Sulp 4000 02 08	0.077
500	Sulp 4000 02 08	0.196
800	Sulp 4000 02 08	0.365
1000	Sulp 4000 02 08	0.452

*serial Dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity	Abs/ppm		
0 - 250 mg/l	0.0003	0.0003 to 0.0005	Pass
250 - 500 mg/l	0.0005	0.0003 to 0.0006	Pass
500 - 800 mg/l	0.0006	0.0004 to 0.0007	Pass
800 - 1000 mg/l	0.0004	0.0004 to 0.0008	Pass

Notes Store between 2 & 8 °c
*Stable on Board 5 days
Barcode ID = 767
* Based on results obtained in our Research Facility

Dr. T. Russell