

Certificate of Performance

Phos 1 Reagent



Product Code 984644  
984646  
Lot Number 8152

Date 15.08.2008


Expiry Date 08.08.2009

Reagent	Lot Number
Phos 1	8152
Phos 2	8105
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.004
0.5	P20 02 08	0.155
1	P20 02 08	0.303
1.5	P20 02 08	0.450
2	P20 02 08	0.593

\*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.006	-0.005 to 0.010	Pass
Sensitivity	Abs/ppm 0.295	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C  
\*Stable on Board 5 days  
Barcode ID = 763  
\* Based on results obtained in our Research Facility

*Tom Russell*

Dr. T. Russell