


Certificate of Performance

Phos 2 Reagent



Product Code 984645
984647
Lot Number 8165

Date 08.09.2008


Expiry Date 29.08.2009

Reagent	Lot Number
Phos 1	8164
Phos 2	8165
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P20 08 08	0.146
1	P20 08 08	0.290
1.5	P20 08 08	0.446
2	P20 08 08	0.584

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.000	-0.005 to 0.010	Pass
Sensitivity	Abs/ppm 0.293	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
*Stable on Board 5 days
Barcode ID = 765
* Based on results obtained in our Research Facility

Tom Russell

Dr. T. Russell