

Certificate of Performance



Chloride Reagent

Product Code 984642
984643
Lot Number 9003


Date 13.01.2009

Expiry Date 02.01.2010

Reagent	Lot Number
Chloride	9003
Test Method	LTP02
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Absorbance
0 ppm		0.004
250 ppm	CL4000 12 08	0.334
500 ppm	CL4000 12 08	0.550
800 ppm	CL4000 12 08	0.721
1000 ppm	CL4000 12 08	0.813

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.004	-0.005 to 0.01	Pass
Sensitivity	Abs/ppm		
	0 - 250 mg/l	0.0010 to 0.0016	Pass
	250 - 500 mg/l	0.0006 to 0.0012	Pass
	500 - 800 mg/l	0.0004 to 0.0010	Pass
	800 - 1000 mg/l	0.0003 to 0.0008	Pass

Notes Store between 2 & 8 °c
*Stable on Board 5 days
Barcode ID = 754
*Based on results obtained in our Research Facility

Tom Russell

Dr. T. Russell