

Certificate of Performance

Sulphate Reagent



Product Code 984648

984649

Lot Number 9055

Date 23.03.2009


Expiry Date 18.03.2010

Reagent	Lot Number
Sulphate	9055
Test Method	LTP01
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.005
250	Sulp 4000 01 09	0.122
500	Sulp 4000 01 09	0.270
800	Sulp 4000 01 09	0.446
1000	Sulp 4000 01 09	0.553

*serial Dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.005	-0.005 to 0.010	Pass
Sensitivity	Abs/ppm		
0 - 250 mg/l	0.0005	0.0003 to 0.0005	Pass
250 - 500 mg/l	0.0006	0.0003 to 0.0006	Pass
500 - 800 mg/l	0.0006	0.0004 to 0.0007	Pass
800 - 1000 mg/l	0.0005	0.0004 to 0.0008	Pass

Notes Store between 2 & 8 °c
*Stable on Board 5 days
Barcode ID = 767
*Based on results obtained in our Research Facility

Dr. T. Russell