

Certificate of Performance

Phos 1 Reagent



Product Code 984644

984646

Lot Number 9077

Date 01.04.2009


Expiry Date 30.03.2010

Reagent	Lot Number
Phos 1	9077
Phos 2	9035
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.004
0.5	P20 01 09	0.157
1	P20 01 09	0.303
1.5	P20 01 09	0.463
2	P20 01 09	0.606

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.005	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.302	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
*Stable on Board 5 days
Barcode ID = 763
* Based on results obtained in our Research Facility

Dr. T. Russell